

Maintenance of Laboratories

AEC/EEE/EM/DTM/GI-14

~~Q. 21~~

ADITYA ENGINEERING COLLEGE
SURAMPALEM - 533 437

:- AEC/EEE/EM/APSM/015

~~CA~~

ADITYA ENGINEERING COLLEGE
SURAMPalem - 533 437

Name of the Equipment :- power factor meter

Equipment ID :- AEC/EEE/EM/AFM/01

Date	Nature of Maintenance done	Maintenance done by	Lab Incharge	Remarks
22/02/22	Meter checked practically	T. Apparao	B	Working is good
08/3/22	Meter cleaned	T. Apparao	B	done
22/3/22	Meter resistance values are checked	T. Apparao	B	Values are good
06/4/22	Meter terminals are checked	T. Apparao	B	They are good
29/4/22	Meter cleaned	T. Apparao	B	done
4/5/22	Meter output is verified practically	T. Apparao	B	Working is good
18/5/22	Meter resistance values are checked	T. Apparao	B	They are good
01/6/22	Meter resistance values are checked	T. Apparao	B	They are good
15/6/22	Meter terminals are checked	T. Apparao	B	They are good
29/6/22	Meter cleaned	T. Apparao	B	done
13/7/22	Meter checked practically	T. Apparao	B	Working is good
27/7/22	Meter resistance values are checked	T. Apparao	B	resistance values are OK
10/8/22	Meter cleaned	T. Apparao	B	done
24/08/22	Meter terminals are checked	T. Apparao	B	They are good

~~2.6~~

Name of the Equipment :- function generator 0-7MHz
Equipment ID :- AEC/EEE/NET/F-6/01-02

Date	Nature of maintenance done	Maintenance done by	Signature	Remarks
12/11/2020	dust cleaned	G.S.P. Rao	Brc	maintenance done
26/11/2020	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
10/12/2020	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
24/12/2020	dust cleaned	G.S.P. Rao	Brc	maintenance done
07/01/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
21/01/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
04/02/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
18/02/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
04/03/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
18/03/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
04/04/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
15/04/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
27/04/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
13/05/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
23/05/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
09/06/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
23/06/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
07/07/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
24/07/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
5/08/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
19/08/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
02/09/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
16/09/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
30/09/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
16/10/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
29/10/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
15/11/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
29/11/2021	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
13/12/2021	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
27/12/2021	dust cleaned	G.S.P. Rao	Brc	maintenance done
10/01/2022	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
24/01/2022	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done
07/02/2022	dust cleaned	G.S.P. Rao	Brc	maintenance done
21/02/2022	Terminal and power cable checked	G.S.P. Rao	Brc	maintenance done
07/03/2022	dust cleaned by using blower	G.S.P. Rao	Brc	maintenance done

PRINCIPAL

Name of the Equipment:-

Cathode Ray Oscilloscope

Equipment ID:-

AEC/ECG/LIC/CRO/1-8

Date	Maintenance done	Person done Name & Sign	Lab incharge Sign	Remarks
8/4/21	Replace the switch button	N. Raj Kumar	Buhary	Maintenance done
29/4/21	Trace line is adjusted	N. Raj Kumar	Buhary	Maintenance done
7/6/21	Channel 1 knob is checked	N. Raj Kumar	Buhary	Maintenance done
28/6/21	CRO probes were soldering	N. Raj Kumar	Buhary	Maintenance done
12/7/21	Focus knobs adjusted	N. Raj Kumar	Buhary	Maintenance done
26/7/21	Volt/div knobs adjusted	N. Raj Kumar	Buhary	Maintenance done
9/8/21	Dust cleaned	N. Raj Kumar	Buhary	Maintenance done
30/8/21	Dust cleaned	N. Raj Kumar	Buhary	Maintenance done
15/9/21	output checked in CRO	N. Raj Kumar	Buhary	Maintenance done
13/9/21	Two channels knobs are checked	N. Raj Kumar	Buhary	Maintenance done
4/10/21	power chord changed	N. Raj Kumar	Buhary	Maintenance done
18/10/21	power chord changed	N. Raj Kumar	Buhary	Maintenance done
3/12/21	Replace the switch button	N. Raj Kumar	Buhary	Maintenance done
24/12/21	Replace the switch button	N. Raj Kumar	Buhary	Maintenance done
7/1/22	focus knobs adjusted	N. Raj Kumar	Buhary	Maintenance done
28/1/22	output checked in CRO	N. Raj Kumar	Buhary	Maintenance done
11/2/22	Volt/div knob adjusted	N. Raj Kumar	Buhary	Maintenance done
25/2/22	Trace line is adjusted	N. Raj Kumar	Buhary	Maintenance done
4/3/22	Trace line is adjusted	N. Raj Kumar	Buhary	Maintenance done
25/3/22	Channel 1 knob is checked	N. Raj Kumar	Buhary	Maintenance done
8/4/22	Dust cleaned	N. Raj Kumar	Buhary	Maintenance done
29/4/22	output checked in CRO	N. Raj Kumar	Buhary	Maintenance done
7/6/22	Two channel knobs are checked	N. Raj Kumar	Buhary	Maintenance done
28/6/22	focus knobs are checked	N. Raj Kumar	BSL	Maintenance done
12/7/22	power chord changed	N. Raj Kumar	BSL	Maintenance done
29/7/22	Dust cleaned	N. Raj Kumar	BSL	Maintenance done
16/8/22	Volt/div knobs adjusted	N. Raj Kumar	BSL	Maintenance done

PRINCIPAL

SVYA ENGINEERING COLLEGE

PALEM - 533 437

Name of the Equipment:

Cathode Ray Oscilloscope

Equipment ID:

AEC/ECE/PDC/CRO/7-12

Date	Maintenance done	Person done Name & Sign	Lab Incharge Sign	Remarks
8/7/21	To Replace the Switch Button	B. V. Mahesh	SA	Maintenance done
15/7/2021	To adjusted in a trace line	B. V. Mahesh	SA	Maintenance done
5/8/21	Cap Probes were soldering	B. V. Mahesh	SA	Maintenance done
12/8/21	Focus knobs adjusted	B. V. Mahesh	SA	Maintenance done
2/9/21	volt/div knobs adjusted	B. V. Mahesh	SA	Maintenance done
30/9/21	dust cleaned	B. V. Mahesh	SA	Maintenance done
7/10/21	output checked in cap	B. V. Mahesh	SA	Maintenance done
21/10/21	Two channels knobs were soldering	B. V. Mahesh	SA	Maintenance done
6/12/21	Power chord changed	B. V. Mahesh	SA	Maintenance done
27/12/21	Replace the Switch Button	B. V. Mahesh	SA	Maintenance done
10/1/22	To adjusted in a trace line	B. V. Mahesh	SA	Maintenance done
31/1/22	CH1 knob is checked	B. V. Mahesh	SA	Maintenance done
7/2/22	Cap Probes were soldering	B. V. Mahesh	SA	Maintenance done
28/2/22	Focus knobs adjusted	B. V. Mahesh	SA	Maintenance done
11/3/22	dust cleaned	B. V. Mahesh	SA	Maintenance done
14/3/22	output checked in cap	B. V. Mahesh	SA	Maintenance done
11/4/22	Focus knobs adjusted	B. V. Mahesh	SA	Maintenance done
25/4/22	CH1 knob is adjusted	B. V. Mahesh	SA	Maintenance done
6/6/22	Cap Probes is checked	B. V. Mahesh	SA	Maintenance done
24/6/22	Dust cleaned	B. V. Mahesh	SA	Maintenance done
4/7/22	VOHS/div knobs adjusted	B. V. Mahesh	SA	Maintenance done
18/7/22	Focus knobs adjusted	B. V. Mahesh	SA	Maintenance done
6/8/22	Cap Probes were soldering	B. V. Mahesh	SA	Maintenance done
13/8/22	output checked in cap	B. V. Mahesh	SA	Maintenance done

PRINCIPAL


Maintenance of Library


Date	Problem/Work
20-08-2021	Xerox Machine Paper Jammed
15-09-2021	System Complaint in Counter
30-09-2021	Tube light Replacement
05-10-2021	Wifi not working
22-10-2021	Fan not working in Counter
11-11-2021	Hard Disk Problem in Counter
20-11-2021	Closed rack Handle Replacement
17-12-2021	System Complaint in Digital Library
31-12-2021	Special Machine Repair

Complainant/Assigned to	Resolved on	Sign	Remark
Mr. Panaksha	23-08-2021	cd 23/8/21	
Mr. Kalesha (System Admin)	15-09-2021	cd 15/9/21	
Mr. Surya (Bldg Supervisor)	30-09-2021	cd 30/9/21	
Mr. Kalesha (System Admin)	05-10-2021	cd 05/10/21	
Mr. Surya (Bldg Supervisor)	23-10-2021	cd 23/10/21	
Mr. Kalesha (System Admin)	11-11-2021	cd 11/11/21	
Mr. Chittibabu	23-11-2021	cd 23/11/21	
Mr. Kalesha (System Admin)	17-12-2021	cd 17/12/21	
Mr. Suryasharaya (A.O.)	03-01-2022	cd 3/1/22	

Date	Problem/Work
4-01-2022	Fan not working in Digital Library
09-02-2022	Xerox Machine Complained Paper Jamed
24-02-2022	Tube light Replacement
11-03-2022	System Compliant in Camter
24-03-2022	Window glass Replacement
06-04-2022	Switch Board Replacemnt
21-04-2022	Wifi not working
10-05-2022	System Compliant in Digital Library
19-05-2022	Fan not working in Counter
8-06-2022	Xerox Machine Paper Jamed
22-06-2022	Net work problem in digital library

Complained/Assigned to	Resolved on	Sign	Remark
Mr Suryav (Building Supervisor)	05-01-2022	cd 21/4/22	
Mr Ramakrishna	10-02-2022	cd 19/4/22	
Mr Suryav Building Supervisor	24-02-2022	cd 24/2/22	
Mr Kaleesha (System Admin)	11-03-2022	cd 11/3/22	
Mr Suryav (Building Supervisor)	29-03-2022	cd 29/3/22	
Mr Veerajw	06-04-2022	cd 21/4/22	
Mr Kaleesha (System Admin)	21-04-2022	cd 21/4/22	
Mr Kaleesha (System Admin)	10-05-2022	cd 10/5/22	
Mr Veerajw	20-05-2022	cd 20/5/22	
Mr Ramakrishna	08-06-2022	cd 31/4/22	
Mr Kaleesha (System Admin)	22-06-2022	cd 22/6/22	

5-07-2022	Xerox Paper - Jamed.
12-07-2022	System Compliant in Center
22-07-2022	Revolving chairs Replacement in Digital Library
 LIBRARIAN Aditya Engineering College SURAMPALEM	

Mr. Rana Krishna	08-07-2022
Mr. Kalesha (System Admin)	12-07-2022
Mr. Surya (Building Supervisor)	22-07-22
 PRINCIPAL Aditya Engineering College SURAMPALEM	

Maintenance of Computers



ADITYA ENGINEERING COLLEGE

An Autonomous Institution
Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956
Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph: 99498 76662

Name Of The Lab : Lab-3

Lab Incharge : K. Gnanakumar raja

Lab Programmer : B. Surya Kumar, Yashwanth

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
29	2/11/21	80	78	2	Smex	my	
30	14/11/21	80	78	2	OX	my	
31	21/11/21	80	78	2	Smex	my	
32	29/11/21	80	78	2	Lab	my	B.N.R. 21/1/22
33	1/12/21	80	78	2	OX	my	
34	18/12/21	80	78	2	Lab	my	
35	23/12/21	80	78	2	OX	my	
36	27/12/21	80	78	2	Lab	my	
37	28/12/21	80	78	2	maye	my	B.N.R. 28/12/21
38	1/1/22	80	78	2	Lab	my	
39	2/1/22	80	78	2	OX	my	
40	3/1/22	80	78	2	Lab	my	
41	4/1/22	80	78	2	Lab	my	
42	5/1/22	80	78	2	Lab	my	
43	6/1/22	80	78	2	OX	my	
44	8/1/22	80	78	2	Lab	my	
45	9/1/22	80	78	2	Smex	my	
46	10/1/22	80	78	2	OX	my	B.N.R. 10/1/22
47	25/1/22	80	78	2	Lab	my	
48	2/2/22	80	78	2	Lab	my	
49	3/2/22	80	78	2	Lab	my	
50	4/2/22	80	78	2	Lab	my	
51	5/2/22	80	78	2	Lab	my	
52	7/2/22	80	78	2	OX	my	
53	9/2/22	80	78	2	Lab	my	
54	18/2/22	80	78	2	OX	my	
55	24/2/22	80	78	2	Smex	my	
56	26/2/22	80	78	2	OX	my	B.N.R. 24/2/22

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph: 99498 76652

Name Of The Lab : Computer Lab-4

Lab Incharge : M. Kalyan Ram

Lab Programmer : P.V. Suresh

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
29	6/7/21	72	71	01	(LAN, Monitor)		
30	9/7/21	72	71	01	(LAN)		
31	13/7/21	72	71	01	LAN		
32	15/7/21	72	70	02	Power cable		b.n.m. 15/7/21
33	19/7/21	72	70	02	power cable		
34	22/7/21	72	70	02	Power cable		
35	26/7/21	72	70	02	Power cable		
36	29/7/21	72	70	02	Power cable		b.n.m. 29/7/21
37	2/8/21	72	70	02	Power cable		
38	4/8/21	72	70	02	Power cable		
39	7/8/21	72	70	02	power cable		
40	10/8/21	72	71	01	Keyboard		b.n.m. 10/8/21
41	17/8/21	72	71	01	Keyboard		
42	23/8/21	72	71	01	Keyboard		
43	27/8/21	72	71	01	Keyboard		
44	31/8/21	72	71	01	Keyboard		b.n.m. 31/8/21
45	1/9/21	72	71	01	lan		
46	4/9/21	72	71	01	lan		
47	9/9/21	72	71	01	lan		
48	14/9/21	72	71	01	lan		b.n.m. 14/9/21
49	16/9/21	72	71	01	lan		
50	18/9/21	72	71	01	lan		
51	21/9/21	72	71	01	lan		
52	25/9/21	72	71	01	lan		
53	29/9/21	72	71	01	lan		b.n.m. 29/9/21
54	1/10/21	72	71	01	lan		
55	5/10/21	72	71	01	lan		
56	8/10/21	72	71	01	lan.		

PRINCIPAL

Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution
Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956
Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G. Dr., Ph 99498 7644

Name Of The Lab : Computer Lab - H

Lab Incharge : M. Kalyan Ram

Lab Programmer : P.V. Suresh, N.S.L. Kalyani

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
29	28/10/21	72	71	01	Monitor	[Signature]	
30	30/10/21	72	71	01	Monitor	[Signature]	
31	2/11/21	72	71	01	Monitor	[Signature]	
32	8/11/21	72	71	01	Monitor	[Signature]	
33	12/11/21	72	71	01	Monitor	[Signature]	
34	15/11/21	72	71	01	Monitor	[Signature]	
35	22/11/21	72	71	01	Monitor	[Signature]	
36	26/11/21	72	71	01	Monitor	[Signature]	
37	30/11/21	72	71	01	Monitor	[Signature]	
38	1/12/21	72	71	01	Monitor	[Signature]	
39	6/12/21	72	72	-	-	[Signature]	
40	9/12/21	72	72	-	-	[Signature]	
41	14/12/21	72	72	-	-	[Signature]	
42	17/12/21	72	72	-	-	[Signature]	
43	22/12/21	72	72	-	-	[Signature]	
44	27/12/21	72	72	-	-	[Signature]	
45	30/12/21	72	72	-	-	[Signature]	
46	3/01/22	72	72	-	-	[Signature]	
47	5/1/22	72	72	-	-	[Signature]	
48	7/1/22	72	70	02	LAN	[Signature]	
49	10/1/22	72	70	02	LAN	[Signature]	
50	12/1/22	72	70	02	LAN	[Signature]	
51	20/1/22	72	70	02	LAN	[Signature]	
52	25/1/22	72	70	02	LAN	[Signature]	
53	29/1/22	72	70	02	LAN	[Signature]	
54	31/1/22	72	70	02	LAN	[Signature]	
55	2/2/22	72	70	02	LAN	[Signature]	
56	7/2/22	72	70	02	LAN	[Signature]	

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph:99498 76667

Name Of The Lab : Computer Lab-4

Lab Incharge : M. Kalyan Ram

Lab Programmer : P.V. Suresh, N.S.K. Kalyani

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
1	10/2/22	72	71	01	Power cable		
2	14/2/22	72	71	01	Power cable		
3	17/2/22	72	71	01	Power cable		
4	22/2/22	72	71	01	power cable		
5	25/2/22	72	71	01	Power cable		
6	28/2/22	72	71	01	Power cable		
7	1/3/22	72	71	01	Power cable		
8	3/3/22	72	71	01	Power cable		
9	5/3/22	72	72	-	-		
10	8/3/22	72	72	-	-		
11	11/3/22	72	72	-	-		
12	15/3/22	72	72	-	-		
13	19/3/22	72	72	-	-		
14	21/3/22	72	72	-	-		
15	23/3/22	72	72	-	-		
16	25/3/22	72	72	-	-		
17	28/3/22	72	72	-	-		
18	30/3/22	72	72	-	-		
19	31/3/22	72	71	01	Monitor		
20	1/4/22	72	71	01	Monitor		
21	4/4/22	72	71	01	Monitor		
22	6/4/22	72	71	01	Monitor		
23	8/4/22	72	71	01	Monitor		
24	11/4/22	72	70	02	lan		
25	13/4/22	72	70	02	LAN		
26	18/4/22	72	70	02	LAN		
27	20/4/22	72	70	02	LAN		
28	22/4/22	72	70	02	LAN		

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE • Permanently Affiliated to JNTUK • Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph:99498 76662

Name Of The Lab : Computer Lab-4

Lab Incharge : M. Kalyan Ram

Lab Programmer : P.V. Suresh, N.S.L. Kalyani

S No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
1	22/4/22	72	71	01	CPU	[Signature]	[Signature]
2	25/4/22	72	71	01	CPU	[Signature]	[Signature]
3	28/4/22	72	71	01	CPU	[Signature]	[Signature]
4	2/5/22	72	71	01	CPU	[Signature]	[Signature]
5	5/5/22	72	71	01	CPU	[Signature]	[Signature]
6	7/5/22	72	71	01	CPU	[Signature]	[Signature]
7	10/5/22	72	71	01	CPU	[Signature]	[Signature]
8	12/5/22	72	71	01	CPU	[Signature]	[Signature]
9	14/5/22	72	71	01	CPU	[Signature]	[Signature]
10	17/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
11	19/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
12	21/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
13	24/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
14	27/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
15	30/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
16	31/5/22	72	70	02	OS, LAN	[Signature]	[Signature]
17	1/6/22	72	70	02	OS, LAN	[Signature]	[Signature]
18	3/6/22	72	70	02	OS, LAN	[Signature]	[Signature]
19	7/6/22	72	70	02	OS, LAN	[Signature]	[Signature]
20	10/6/22	72	71	01	LAN	[Signature]	[Signature]
21	14/6/22	72	71	01	LAN	[Signature]	[Signature]
22	17/6/22	72	71	01	LAN	[Signature]	[Signature]
23	20/6/22	72	71	01	LAN	[Signature]	[Signature]
24	23/6/22	72	71	01	LAN	[Signature]	[Signature]
25	25/6/22	72	71	01	LAN	[Signature]	[Signature]
26	27/6/22	72	71	01	LAN	[Signature]	[Signature]
27	29/6/22	72	71	01	LAN	[Signature]	[Signature]
28	30/6/22	72	71	01	LAN	[Signature]	[Signature]

S.L.

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E. G. Dt. Ph: 99-498 7666

Name Of The Lab : Computer lab - H

Lab Incharge : M. Kalyan Ram

Lab Programmer : P.V. Suresh, N.S.L. Kalyani

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
29	1/7/22	72	70	02	CPU, LAN		
30	2/7/22	72	70	02	CPU, LAN		
31	4/7/22	72	70	02	CPU, LAN		
32	6/7/22	72	70	02	CPU, LAN		
33	8/7/22	72	70	02	CPU, LAN		
34	11/7/22	72	70	02	CPU, LAN		
35	13/7/22	72	70	02	CPU, LAN		
36	14/7/22	72	70	02	CPU, LAN		
37	15/7/22	72	70	02	CPU, LAN		
38	16/7/22	72	70	02	CPU, LAN		
39	18/7/22	72	68	04	CPU, OS, LAN		
40	19/7/22	72	68	04	CPU, OS, LAN		
41	20/7/22	72	68	04	CPU, OS, LAN		
42	21/7/22	72	68	04	CPU, OS, LAN		
43	23/7/22	72	68	04	CPU, OS, LAN		
44	25/7/22	72	68	04	CPU, OS, LAN		
45	27/7/22	72	68	04	CPU, OS, LAN		
46	28/7/22	72	68	04	CPU, OS, LAN		
47	30/7/22	72	68	04	CPU, OS, LAN		
48	2/8/22	72	68	04	CPU, OS, LAN		
49	4/8/22	72	68	04	CPU, OS, LAN		
50	5/8/22	72	70	02	OS, LAN		
51	6/8/22	72	70	02	OS, LAN		
52	8/8/22	72	70	02	OS, LAN		
53	10/8/22	72	70	02	OS, LAN		
54	12/8/22	72	70	02	OS, LAN		
55	13/8/22	72	70	02	OS, LAN		
56	15/8/22	72	70	02	OS, LAN		

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution
Approved by AICTE • Permanently Affiliated to JNTUK • Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956
Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph:99498 76662

Name Of The Lab : LAB-3

Lab Incharge : DR. M. Venkatesh

Lab Programmer : B.surya kiran, y. priyanka, y. Amrutha

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
1	2/3/22	80	78	2	SMPS	u/	
2	3/3/22	80	78	2	OS	u/	
3	4/3/22	80	79	1	OS	u/	
4	5/3/22	80	79	1	MOUSE	u/	B.M.R. 5/3/22
5	7/3/22	80	79	1	MOUSE	u/	
6	8/3/22	80	80	NILL	Keyboard	u/	
7	9/3/22	80	80	NILL	Mat -	u/	
8	10/3/22	80	78	2	MOUSE	u/	
9	11/3/22	80	78	2	OS	u/	
10	12/3/22	80	79	1	OS	u/	
11	14/3/22	80	80	NILL	-	u/	B.M.R. 14/3/22
12	15/3/22	80	80	NILL	-	u/	
13	16/3/22	80	80	NILL	-	u/	
14	17/3/22	80	77	3	OS, LAN	u/	
15	19/3/22	80	78	2	LAN	u/	
16	21/3/22	80	78	2	LAN	u/	
17	22/3/22	80	78	2	LAN	u/	
18	23/3/22	80	78	2	OS	u/	B.M.R. 23/3/22
19	24/3/22	80	79	1	OS	u/	
20	25/3/22	80	79	1	LAN	u/	
21	26/3/22	80	79	1	MOUSE	u/	
22	28/3/22	80	80	NILL	-	u/	
23	29/3/22	80	80	NILL	-	u/	
24	30/3/22	80	80	NILL	-	u/	
25	31/3/22	80	78	2	Keyboard	u/	B.M.R. 31/3/22
26	1/4/22	80	78	2	Keyboard	u/	
27	4/4/22	80	78	2	OS	u/	
28	6/4/22	80	78	2	OS	u/	B.M.R. 6/4/22

PRINCIPAL
Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G. Dt., Ph:99498 76662

Name Of The Lab : LAB-3

Lab Incharge : DR.M.Venkatesh

Lab Programmer : B.Surya Kiran, Y.priyanka, Y.Amratha

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
29	8/4/22	80	78	2	OS	✓	
30	9/4/22	80	78	2	LAN	✓	
31	11/4/22	80	78	2	Mouse	✓	
32	12/4/22	80	78	2	SMPS	✓	
33	13/4/22	80	79	1	LAN	✓	B.V.R. 13/4/22
34	18/4/22	80	79	1	LAN	✓	
35	19/4/22	80	80	NILL	-	✓	
36	20/4/22	80	78	2	OS	✓	
37	21/4/22	80	78	2	OS	✓	
38	22/4/22	80	78	2	OS	✓	
39	23/4/22	80	78	2	LAN	✓	B.V.R. 23/4/22
40	25/4/22	80	79	1	Mouse	✓	
41	26/4/22	80	79	1	Keyboard	✓	
42	27/4/22	80	79	1	OS	✓	
43	28/4/22	80	79	1	OS	✓	
44	29/4/22	80	79	1	LAN	✓	
45	30/4/22	80	80	NILL	-	✓	B.V.R. 30/4/22
46	2/5/22	80	80	NILL	-	✓	
47	4/5/22	80	80	NILL	-	✓	
48	5/5/22	80	78	2	Mouse	✓	
49	6/5/22	80	78	2	Keyboard	✓	
50	7/5/22	80	78	2	Mouse	✓	B.V.R. 7/5/22
51	9/5/22	80	78	2	Mouse	✓	
52	10/5/22	80	79	1	LAN	✓	
53	11/5/22	80	79	1	LAN	✓	
54	12/5/22	80	79	1	LAN	✓	
55	13/5/22	80	78	2	OS	✓	B.V.R. 13/5/22
56	14/5/22	80	79	1	OS	✓	B.V.R. 14/5/22

PRINCIPAL

Aditya Engineering College
SURAMPALAM



ADITYA ENGINEERING COLLEGE

An Autonomous Institution
Approved by AICTE - Permanently Affiliated to JNTUK - Accredited by NAAC with 'A' Grade
Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956
Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt. Ph: 99498 76662

Name Of The Lab : Lab-3

Lab Incharge : Dr. M. Venkatesh

Lab Programmer : B. Surya Kumar, K. Priyanka, Y. Anuradha

S.No	Date	No Of Systems	No Of Systems Working	No Of Systems Not Working	Complaints if any in the lab	Signature (Incharge)	Signature(HOD)
1	16/5/22	80	78	02	Lan on	u	
2	17/5/22	80	78	02	CPU	u	
3	18/5/22	80	78	02	CPU	u	
4	19/5/22	80	78	02	Lan on	u	
5	20/5/22	80	78	02	Smpr	u	
6	21/5/22	80	78	02	Lan	u	
7	23/5/22	80	78	02	CPU	u	
8	24/5/22	80	78	02	CPU	u	
9	26/5/22	80	78	02	Smpr	u	
10	27/5/22	80	78	02	Lan	u	
11	28/5/22	80	78	02	Lan	u	
12	1/6/22	80	78	02	Lan	u	
13	2/6/22	80	78	02	Lan	u	
14	3/6/22	80	78	02	on	u	
15	4/6/22	80	78	02	on	u	
16	6/6/22	80	78	02	CPU	u	
17	7/6/22	80	78	02	Smpr	u	
18	8/6/22	80	78	02	Matlab	u	
19	9/6/22	80	78	02	Smpr	u	
20	10/6/22	80	78	02	Smpr	u	
21	11/6/22	80	78	02	CPU	u	
22	13/6/22	80	78	02	CPU	u	
23	15/6/22	80	78	02	Smpr	u	
24	16/6/22	80	78	02	Lan	u	
25	17/6/22	80	78	02	Lan	u	
26	18/6/22	80	78	02	on	u	
27	19/6/22	80	78	02	on	u	
28	20/6/22	80	78	02	on	u	

ADITYA ENGINEERING COLLEGE

An Autonomous Institution

Approved by AICTE • Permanently Affiliated to JNTUK • Accredited by NAAC with 'A' Grade

Recognised by UGC under sections 2(f) and 12(B) of UGC Act, 1956

Aditya Nagar, ADB Road, Surampalem - 533437, Near Kakinada, E.G.Dt., Ph:99498 76662

Name Of The Lab : Lab-3

Lab Incharge : Dr. M. Venkatesh

Lab Programmer : B. Swaga KNCU, K. Priyanka

[illegible]

PRINCIPAL
Aditya Engineering College
SURAMPAL EM